

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed Diluent**
Article number: **QRR 010931**
LOT number: **2106109817**
Epiry date: **2024-03**

LOT release date: **2021-04-28**

| Test | Specification | Measured value |
|----------------------|---------------|----------------|
| Conductivity (mS) | 18.0 – 19.0 | 18.7 |
| Osmolality (mOsm/kg) | 331 – 341 | 334 |
| pH at 25°C | 5.55 – 5.75 | 5.60 |

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed:

Date: *18 May 2021*

QARA Department




