

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAG  
The Netherlands

T +31 229 291 129  
www.rrmechatronics.com

## Certificate of analysis

Product name: **Starsed Diluent**  
Article number: **QRR 010931**  
LOT number: **2216609812**  
Expiry date: **2025-06**

LOT release date: **2022-07-07**

Test	Specification	Measured value
Conductivity (mS)	18.0 – 19.0	18.9
Osmolality (mOsm/kg)	331 – 341	334
pH at 25°C	5.55 – 5.75	5.60

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).  
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed:

Date: 18-jul-2022

QARA Department

