

Certificate of analysis

Product name: **Starsed Diluent**
Article number: **QRR 010931**
LOT number: 2314109812
Expiry date: 2026-05

LOT release date: 2023-06-02

Test	Specification	Measured value
Conductivity (mS)	18.0 – 19.0	18.7
Osmolality (mOsm/kg)	331 – 341	333
pH at 25°C	5.55 – 5.75	5.64

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).
For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed:



Date:

20-Jun-2023

QARA Department





Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No QRR010931
DILUENT

Lot No 2314109812

Release Date 02-Jun-2023

Test	Specification	Actual value
Conductivity (mS)	18.0 - 19.0	18.7
Osmolality (mOsm/kg)	331 - 341	333
pH at 25°C	5.55 - 5.75	5.64

Use Before Date: 2026-05

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

A handwritten signature in black ink that reads "James T Ethler".

Jamie Ethler
Vice President Global Quality

