

RR Mechatronics Manufacturing B.V.

P.O. Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starsted Disinfectant**
Article number: **QRR 010947**
LOT number: 2408909825
Expiry date: 2026-03

LOT release date: 2024-05-27


Test	Specification	Measured value
Active chlorine	0.9 – 1.1 %	1.0%

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).
For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed: 
Date: 1 jul 2024
QARA Department

