

Certificate of analysis

Product name: **Starrsed Rinse solution**
Article number: **QRR 010934**
LOT number: **1901402830**
Expiry date: **2021-01**

LOT release date: **2019-03-07**

Test	Specification	Measured value
Conductivity (mS)	1020 - 1060	1046
Performance test*	passes test	passes test
pH at 25°C	7.10 - 7.60	7.54
Total Bacterial Growth (CFU/100ml)	Max. 5	4

*) All parameters except mid cells: Test/Ref ration > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com


Konstantin Artz

Manager QA/RA

