RR Mechatronics
Masters of Measurement

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

# Certificate of analysis

Product name: **Starrsed Rinse solution**
Article number: **QRR 010934**
LOT number: **2015902832**
Expiry date: **2022-06**

LOT release date: **2020-07-03**

| Test | Specification | Measured value |
|---|---|---|
| Conductivity (mS) | 1020 - 1060 | 1057 |
| Performance test* | passes test | passes test |
| pH at 25°C | 7.10 - 7.60 | 7.60 |
| Total Bacterial Growth (CFU/100ml) | Max. 5 | 1 |

*) All parameters except mid cells: Test/Ref ration > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed:

Date: 01 sep 2020

QARA Department



Conditions: By the ORGALIME GENERAL CONDITIONS S2012 of March 2012. Any other conditions are herewith explicitly rejected by us.
Bank: ABN AMRO, Account number: 0403.537.037, Swift-address: ABNANL2A, IBAN: NL60 ABNA 0403 537 037
VAT number: NL 8557 65 392 B01 CoC Number 64657353