

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed Rinse solution**
Article number: **QRR 010934**
LOT number: **2132609814**
Expiry date: **2023-11**

LOT release date: **2021-12-06**

Test	Specification	Measured value
Conductivity (mS)	1020 - 1060	1038
Performance test*	passes test	passes test
pH at 25°C	7.10 - 7.60	7.47
Total Bacterial Growth (CFU/100ml)	Max. 5	1

*) All parameters except mid cells: Test/Ref ration > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed:



Date: 18-Jan-2022

QARA Department





Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chem

Certificate of Analysis

Product No QRR010934
RINSE
Lot No 2132609814
Release Date 06-Dec-2021

Test	Specification	Actual value
Conductivity (uS)	1020 - 1060	1038
Performance Test *	passes test	passes test
pH at 25°C	7.10 - 7.60	7.47
Total Bacterial Growth (CFU/100ml)	max. 5	1

*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

Use Before Date: 2023-11

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

A handwritten signature in black ink that reads "Jamie Ethler".

Jamie Ethler
Vice President Global Quality

