

Certificate of analysis

Product name: **Starsed Rinse solution**
Article number: **QRR 010934**
LOT number: **22011009811**
Expiry date: **2024-01**

LOT release date: **2022-01-24**

| Test | Specification | Measured value |
|------------------------------------|---------------|----------------|
| Conductivity (mS) | 1020 - 1060 | 1022 |
| Performance test* | passes test | passes test |
| pH at 25°C | 7.10 - 7.60 | 7.58 |
| Total Bacterial Growth (CFU/100ml) | Max. 5 | 0 |

*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed:



Date:

28- Feb - 2022

QARA Department



2/24/22, 9:45 AM

Product#:QRR010934 - Lot#:2201109811



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No QRR010934
RINSE
Lot No 2201109811
Release Date 24-Jan-2022

| Test | Specification | Actual value |
|------------------------------------|---------------|--------------|
| Conductivity (uS) | 1020 - 1060 | 1022 |
| Performance Test * | passes test | passes test |
| pH at 25°C | 7.10 - 7.60 | 7.58 |
| Total Bacterial Growth (CFU/100ml) | max. 5 | 0 |

*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

Use Before Date: 2024-01

For Laboratory, Research or Manufacturing Use.
This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com
The information in this document is property of Avantor Performance Materials Poland S.A.
Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:
For information on Use Descriptors please contact us by mail to
JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

A handwritten signature in black ink that reads "Jamie Ethler".
Jamie Ethler
Vice President Global Quality

