

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAG  
The Netherlands

T +31 229 291 129  
www.rrmechatronics.com

## Certificate of analysis

Product name: **Starsed Rinse solution**  
Article number: **QRR 010934**  
LOT number: 2227109811  
Expiry date: 2024-09

LOT release date: 2022-10-10

| Test                               | Specification | Measured value |
|------------------------------------|---------------|----------------|
| Conductivity (mS)                  | 1020 - 1060   | 1035           |
| Performance test*                  | passes test   | passes test    |
| pH at 25°C                         | 7.10 - 7.60   | 7.53           |
| Total Bacterial Growth (CFU/100ml) | Max. 5        | 0              |

\*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed:



Date:

24-Jan-2023

QARA Department







Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

## Certificate of Analysis

**Product No** QRR010934  
RINSE

**Lot No** 2227109811

**Release Date** 10-Oct-2022

| Test                               | Specification | Actual value |
|------------------------------------|---------------|--------------|
| Conductivity (uS)                  | 1020 - 1060   | 1035         |
| Performance Test *                 | passes test   | passes test  |
| pH at 25°C                         | 7.10 - 7.60   | 7.53         |
| Total Bacterial Growth (CFU/100ml) | max. 5        | 0            |

\*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

Use Before Date: 2024-09

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at [avantor.emea@avantormaterials.com](mailto:avantor.emea@avantormaterials.com)

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiago 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker\_REACH\_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

A handwritten signature in black ink that reads "Jamie Ethler".

Jamie Ethler  
Vice President Global Quality

