

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAG  
The Netherlands

T +31 229 291 129  
[www.rrmechatronics.com](http://www.rrmechatronics.com)

## Certificate of analysis

Product name: **Starrsed Rinse solution**  
Article number: **QRR 010934**  
LOT number: 2314809812  
Expiry date: 2025-05

LOT release date: 2023-06-26

| Test                               | Specification | Measured value |
|------------------------------------|---------------|----------------|
| Conductivity (mS)                  | 1020 - 1060   | 1030           |
| Performance test*                  | passes test   | passes test    |
| pH at 25°C                         | 7.10 - 7.60   | 7.57           |
| Total Bacterial Growth (CFU/100ml) | Max. 5        | 0              |

\*) All parameters except mid cells: Test/Ref ratio > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed:



Date:

07. aug - 2023

QARA Department

