



Certificate of analysis

Product name: **Starrsed Saline**
Article number: **QRR 010933**
LOT number: **2104709810**
Expiry date: **2024-02**

LOT release date: **2021-04-19**

Test	Specification	Measured value
Conductivity (mS)	15.20 - 16.20	15.86
Osmolality (mOsm/kg)	280 - 292	289
pH at 25°C	7.20 - 7.30	7.26
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed:

Date: 09-08-2021

QARA Department

