

Certificate of analysis

Product name: **Starrsed Saline**
Article number: **QRR 010933**
LOT number: **2201909814**
Expiry date: **2025-01**

LOT release date: **2022-01-28**

Test	Specification	Measured value
Conductivity (mS)	15.20 - 16.20	16.02
Osmolality (mOsm/kg)	280 - 292	291
pH at 25°C	7.20 - 7.30	7.22
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed:



Date:

28 Feb. 2022

QARA Department





Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No QRR010933
SALINE
Lot No 2201909814
Release Date 28-Jan-2022

Test	Specification	Actual value
Conductivity (mS)	15.20 - 16.20	16.02
Osmolality (mOsm/kg)	280 - 292	291
pH at 25°C	7.20 - 7.30	7.22
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	max. 5	0

Use Before Date: 2025-01


For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.
Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel: +48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:
For information on Use Descriptors please contact us by mail to
JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.



Jamie Ethier
Vice President Global Quality

