

RR Mechatronics Manufacturing B.V.

 P.O.Box 225
 1620 AE HOORN
 The Netherlands

 De Corantijn 13
 1689 AN ZWAAG
 The Netherlands

 T +31 229 291 129
 www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed Saline**
 Article number: **QRR 010933**
 LOT number: **2222809827**
 Expiry date: **2025-08**

LOT release date: **2022-08-29**

Test	Specification	Measured value
Conductivity (mS)	15.20 - 16.20	16.04
Osmolality (mOsm/kg)	280 - 292	287
pH at 25°C	7.20 - 7.30	7.23
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
 For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed: 

Date: *26 sep 2022*

QARA Department





Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No QRR010933
SALINE

Lot No 2222809827

Release Date 29-Aug-2022

Test	Specification	Actual value
Conductivity (mS)	15.20 - 16.20	16.04
Osmolality (mOsm/kg)	280 - 292	287
pH at 25°C	7.20 - 7.30	7.23
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	max. 5	0

Use Before Date: 2025-08

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this cofa is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

A handwritten signature in black ink that reads "Jamie Ethler".

Jamie Ethler
Vice President Global Quality

