

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starsed Saline**
Article number: **QRR 010933**
LOT number: 2308609810
Expiry date: 2026-03

LOT release date: 2023-04-07

Test	Specification	Measured value
Conductivity (mS)	15.20 - 16.20	15.67
Osmolality (mOsm/kg)	280 - 292	283
pH at 25°C	7.20 - 7.30	7.21
Presence of Tween	passes test	passes test
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).
For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed:



Date:

18-apr-2023

QARA Department

