

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed Control Level A**
Article number: **QRR 049002**

LOT number: **QCA945AA28**
Expiry date: **2022-08-22**

LOT release date: **2021-06-14**

| | Specification | Validated value |
|-------------|---------------|-----------------|
| Value range | 30-60 | 30-50 |
| Mean value | 40-50 | 40 |

The following information is provided by our OEM-manufacturer

1. No microbial contamination was recorded, as per quality control procedures
2. Fill volumes were verified to meet all specifications
3. Each human donor unit used to manufacture this product was tested and found non-reactive at the donor level per current applicable FDA requirements using FDA-accepted methods including testing for HIV-1/HIV-2 Antibody, HBsAg, HCV Antibody, HIV-1 and HCV by a nucleic acid test (NAT)
4. Batch uniformity studies verified product specifications were met

This lot of product has been controlled and approved for distribution.

For technical information and instructions for use regarding this product, refer to the Package insert.


For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed by:

Date: 01-07-2021

QARA Department




COA_Starrsed_Control_A_QCA945AA28.docx

page 1/1

Conditions: By the ORGALIME GENERAL CONDITIONS S2012 of March 2012. Any other conditions are herewith explicitly rejected by us.

Bank: ABN AMRO, Account number: 0403.537.037, Swift-address: ABNANL2A, IBAN: NL60 ABNA 0403 537 037

VAT number: NL 8557 65 392 B01 CoC Number 64657353