RR Mechatronics
Masters of Measurement

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed X-Clean**
Article number: **QRR 010946**

LOT number: **1703602824**
Expiry date: **2019-02**

LOT release date: **2017-03-14**

Test	Specification	Measured value
Osmolality (mOsm/kg)	395 - 415	407
Performance test	passes test	passes test
pH at 25°C	8.5 - 9.1	8.6
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Konstantin Artz

Manager QA/RA

Conditions: By the ORGALIME GENERAL CONDITIONS S2012 of March 2012. Any other conditions are herewith explicitly rejected by us.
Bank: ABN AMRO, Account number: 0403.537.037, Swift-address: ABNANL2A, IBAN: NL60 ABNA 0403 537 037
VAT number: NL 8557 65 392 B01 CoC Number 64657353