



Certificate of analysis

Product name: **Starrsed X-Clean**
Article number: **QRR 010946**

LOT number: **1828902822**
Expiry date: **2020-10**

LOT release date: **2018-11-07**

Test	Specification	Measured value
Osmolality (mOsm/kg)	395 - 415	405
Performance test	passes test	passes test
pH at 25°C	8.5 - 9.1	8.8
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Konstantin Artz

Manager QA/RA

