

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAAG  
The Netherlands

T +31 229 291 129  
[www.rrmechatronics.com](http://www.rrmechatronics.com)

## Certificate of analysis

Product name: **Starrsed X-Clean**  
Article number: **QRR 010946**

LOT number: **2134009817**  
Expiry date: **2023-12**

LOT release date: **2022-01-28**

Test	Specification	Measured value
Osmolality (mOsm/kg)	395 - 415	409
Performance test	passes test	passes test
pH at 25°C	8.5 - 9.1	8.6
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).  
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed:



Date:

28. feb - 2022

QARA Department







Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

### Certificate of Analysis

Product No      QRR010946  
 X-CLEAN

Lot No            2134009817

Release Date    28-Jan-2022

Test	Specification	Actual value
Osmolality (mOsm/kg)	395 - 415	409
Performance Test	passes test	passes test
pH at 25°C	8.5 - 9.1	8.6
Total Bacterial Growth (CFU/100ml)	max. 5	0

pH is measured directly after production. In course of time the actual value will decrease. This will not influence the performance of the reagent.

Use Before Date: 2023-12

For Laboratory, Research or Manufacturing Use.  
 This document is automatically generated.  
 For questions on this cofa, please contact Customer Service at [avea@avantormaterials.com](mailto:avea@avantormaterials.com)  
 The information in this document is property of Avantor Performance Materials Poland S.A.  
 Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel: +48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:  
 For information on Use Descriptors please contact us by mail to  
[JTBaker\\_REACH\\_USERS@avantormaterials.com](mailto:JTBaker_REACH_USERS@avantormaterials.com) indicating the JT Baker product number you purchase.



Jamie Ethier  
 Vice President Global Quality





