

RR Mechatronics Manufacturing B.V.

 P.O.Box 225  
 1620 AE HOORN  
 The Netherlands

 De Corantijn 13  
 1689 AN ZWAAG  
 The Netherlands

 T +31 229 291 129  
 www.rrmechatronics.com

## Certificate of analysis

Product name: **Starrsed X-Clean**  
 Article number: **QRR 010946**  
 LOT number: **2222209815**  
 Expiry date: **2024-08**

LOT release date: **2022-09-15**

| Test                               | Specification | Measured value |
|------------------------------------|---------------|----------------|
| Osmolality (mOsm/kg)               | 395 - 415     | 408            |
| Performance test                   | passes test   | passes test    |
| pH at 25°C                         | 8.5 - 9.1     | 8.6            |
| Total Bacterial Growth (CFU/100ml) | Max. 5        | 0              |

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).  
 For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed: 

Date: *18-oct-2022*

QARA Department

