

RR Mechatronics Manufacturing B.V.

 P.O.Box 225
 1620 AE HOORN
 The Netherlands

 De Corantijn 13
 1689 AN ZWAAAG
 The Netherlands

 T +31 229 291 129
 www.rrmechatronics.com

Certificate of analysis

Product name: **Starrsed X-Clean**
 Article number: **QRR 010946**
 LOT number: 2408009812
 Expiry date: 2026-03

LOT release date: 2024-03-27

Test	Specification	Measured value
Osmolality (mOsm/kg)	395 - 415	396
Performance test	passes test	passes test
pH at 25°C	8.5 - 9.1	8.8
Total Bacterial Growth (CFU/100ml)	Max. 5	0

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed: 

Date: 13-may-2024

QARA Department

