

# PRODUCT BULLETIN

Type: Field Service

Bulletin Number SB 2021011

## Subject: Volume change Oxy ISO

Date: October 2021

From: Support department

Number of pages: 1

Phone: +31 229 291 129

### Instruments:

All Lorrca instruments.

### Modification:

Oxy ISO QRR030905 will be delivered in 5 ml vials again.

### Reason:

Elon ISO will still be supplied in 2.5 ml vials until further notice.

Due to the Covid19 pandemic, we are still facing a shortage and related price increase of plastic laboratory supplies. Elon ISO is still affected by this shortage. To continue operation with the Lorrca instruments we will continue to deliver this reagent temporarily in alternative/smaller vials without price change.

**Impact on**            **Quality / validity of results:** No  
                             **Reliability of the instrument:** No  
                             **Safety of the instrument:** No

### Effective from:

Oxy ISO batch 2205066B

### Compatibility:

N/A

### Is field modification possible?

N/A

### Order information:

Oxy ISO QRR030905

### Notes on stock and obsolescence:

N/A

### Additional information:

N/A