

Certificate of analysis

Product name: **Elon ISO**
Article number: **QRR 030901**

LOT number: **2103648B**
Expiry date: **2021-03-09**

LOT release date: **2020-03-09**

Test	Specification	Measured value
Osmolality (mOsm/kg)	284 - 304	296
pH at 25°C	7.35 - 7.45	7.38
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	29

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rremechatronics.com

Signed by:

Date: *16-may-2020*

QARA Department

