

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Elon ISO**
Article number: **QRR 030901**
LOT number: 2402186B
Expiry date: 2024-02-01

LOT release date: 2023-02-01

Test	Specification	Measured value
Osmolality (mOsm/kg)	284 - 304	291
pH at 25°C	7.35 - 7.45	7,36
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	28,24


This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed by: 

Date: 06 - mar - 2023

QARA Department



ANALYSECERTIFICAAT

Afgifte, Meppel 1 maart 2023

Rabroekenweg 20, 7942 JE Meppel
P.O. Box 37, 7940 AA Meppel
The Netherlands

T. +31 (0)522-26 87 00
☎ +31 (0)6-53 859 900
E. boombv@boomlab.nl

www.boomlab.nl

Naam : Elon ISO
Boom artikelnummer : 80902361
Mechatronics artikelnummer : QRR030901
Batchnummer : 2402186B

Bewaarcondities : 4 - 30 °C, goed afgesloten
Bereidingsdatum : 01-02-2023
Vervaldatum : 01-02-2024

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	284 - 304	291
pH meting (37°C)	geen	7,35 – 7,45	7,36
viscositeit (37°C)	mPas·s	27 - 33	28,24
Aantal keren geautoclaveerd		max 1	1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.



