

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rmechatronics.com

Certificate of analysis

Product name: **Oxy ISO**
Article number: **QRR 030905 (2.5 ml vials)**
LOT number: **2202644B**
Expiry date: **2022-02-17**

LOT release date: **2021-02-17**

Test	Specification	Measured value
Osmolality (mOsm/kg)	282 - 286	285
pH at 25°C	7.35 - 7.45	7.41
Viscosity at 37.0 ± 0.5 °C (cP)	28 - 30	29.26

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at qara@rrmechatronics.com

Signed by:



Date:

23-mar-2021

QARA Department



ANALYSECERTIFICAAT

Afgifte, Meppel 23 maart 2021

Rabroekenweg 20, 7942 JE Meppel
P.O. Box 37, 7940 AA Meppel
The Netherlands

+31 (0)522-26 87 00
+31 (0)6-53 859 900
boombv@boomlab.nl

www.boomlab.nl

Naam : Oxy ISO
Boom artikelnummer : 80902644
Mechatronics artikelnummer : QRR030905-B01
Batchnummer : 2202644B

Bewaarcondities : 4 - 30 °C, goed afgesloten
Bereidingsdatum : 17 februari 2021
Vervaldatum : 17 februari 2022

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	282 - 286	285
pH meting (37°C)	geen	7,35 – 7,45	7,41
viscositeit (37°C)	mPas·s	28 - 30	28.59

Aantal keren geautoclaveerd : max 1 : 1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.