



RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Osmo LOW**
Article number: **QRR 030910**
LOT number: **2111294B**
Expiry date: **2021-11-24**

LOT release date: **2020-11-24**

Test	Specification	Measured value
Osmolality (mOsm/kg)	≤ 55	47
pH at 25°C	7.25-7.45	7.35
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	31.84

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed by:

Date: 15-Dec-2020

QARA Department

