

## Certificate of analysis

Product name: **Osmo LOW**  
Article number: **QRR 030910**  
LOT number: **2212077B**  
Expiry date: **2022-12-24**

LOT release date: **2021-12-24**

Test	Specification	Measured value
Osmolality (mOsm/kg)	≤ 55	39
pH at 25°C	7.25-7.45	7.34
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	28.09

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).  
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed by:



Date: 10-Jan-2022

QARA Department



# ANALYSECERTIFICAAT

Afgifte, Meppel 3 januari 2022

Rabroekenweg 20, 7942 JE Meppel  
P.O. Box 37, 7940 AA Meppel  
The Netherlands

T. +31 (0)522-26 87 00  
☎ +31 (0)6-53 859 900  
E. boombv@boomlab.nl

[www.boomlab.nl](http://www.boomlab.nl)

Naam : Osmo LOW  
Boom artikelnummer : 80902656  
Mechatronics artikelnummer : QRR030910 B01  
Batchnummer : 2212077B

Bewaarcondities : 4 - 30 °C, goed afgesloten  
Bereidingsdatum : 24-12-2021  
Vervaldatum : 24-12-2022

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	≤ 55	39
pH meting (37°C)	geen	7,25 – 7,45	7.34
viscositeit (37°C)	mPas-s	27 - 33	28.09
Aantal keren geautoclaveerd		max 1	1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.

