

Certificate of analysis

Product name: **Osmo LOW**
Article number: **QRR 030910**
LOT number: 2403333B
Expiry date: 2024-03-07

LOT release date: 2023-03-07

Test	Specification	Measured value
Osmolality (mOsm/kg)	≤ 55	49
pH at 25°C	7.25-7.45	7,26
Viscosity at 37.0 \pm 0.5 °C (cP)	27.0 - 33.0	29,26

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed by:



Date: 13-mar-2023

QARA Department



ANALYSECERTIFICAAT

Afgifte, Meppel 8 maart 2023

Rabroekenweg 20, 7942 JE Meppel
P.O. Box 37, 7940 AA Meppel
The Netherlands

T. +31 (0)522-26 87 00
☎ +31 (0)6-53 859 900
E. boombv@boomlab.nl

www.boomlab.nl

Naam : Osmo LOW
Boom artikelnummer : 80902656
Mechatronics artikelnummer : QRR030910 B01
Batchnummer : 2403333B

Bewaarcondities : 4 - 30 °C, goed afgesloten
Bereidingsdatum : 7 maart 2023
Vervaldatum : 7 maart 2024

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	≤ 55	49
pH meting (37°C)	geen	7,25 – 7,45	7,26
viscositeit (37°C)	mPas·s	27 - 33	29,26

Aantal keren geautoclaveerd : max 1 1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.



