

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAG  
The Netherlands

T +31 229 291 129  
[www.rrmechatronics.com](http://www.rrmechatronics.com)

## Certificate of analysis

Product name: **Osmo LOW**  
Article number: **QRR 030910**  
LOT number: 2406025B  
Expiry date: 2024-06-19

LOT release date: 2023-06-19

Test	Specification	Measured value
Osmolality (mOsm/kg)	$\leq 55$	47
pH at 25°C	7.25-7.45	7.26
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	27.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).  
For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [qara@rrmechatronics.com](mailto:qara@rrmechatronics.com)

Signed by:



Date: 17-jul - 2023

QARA Department

