

RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Osmo LOW**
Article number: **QRR 030910**
LOT number: 2507621B
Expiry date: 2025-07-10

LOT release date: 2024-07-10

Test	Specification	Measured value
Osmolality (mOsm/kg)	≤ 55	47
pH at 25°C	7.25-7.45	7.25
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	28.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).
For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed by: 

Date: 10 Jul 2024

QARA Department



ANALYSECERTIFICAAT

Afgifte, Meppel 10 juli 2024

Rabroekenweg 20, 7942 JE Meppel
P.O. Box 37, 7940 AA Meppel
The Netherlands

T: +31 (0)522-26 87 00
F: +31 (0)6-53 859 900
E: boombv@boomlab.nl

www.boomlab.nl

Naam : Osmo LOW
Boom artikelnummer : 80902362
Mechatronics artikelnummer : QRR030910 C01
Batchnummer : 2507621B

Bewaarcondities : 4 - 30 °C, goed afgesloten
Bereidingsdatum : 10-07-2024
Vervaldatum : 10-07-2025

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	≤ 55	47
pH meting (37°C)	geen	7,25 – 7,45	7,25
viscositeit (37°C)	mPas·s	27 - 33	28,95
Aantal keren geautoclaveerd		max 1	1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.

