



RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAAG  
The Netherlands

T +31 229 291 129  
[www.rrmechatronics.com](http://www.rrmechatronics.com)

## Certificate of analysis

Product name: **Osmo HIGH**  
Article number: **QRR 030930**

LOT number: **2302041B**  
Expiry date: **2023-02-25**

LOT release date: **2022-02-25**

Test	Specification	Measured value
Osmolality (mOsm/kg)	764 - 804	768
pH at 25°C	7.35 - 7.45	7.41
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	27.49

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Technical Data Sheet (TDS).  
For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed:

Date:

20- Feb - 2022

QARA Department





# ANALYSECERTIFICAAT

Afgifte, Meppel 25 februari 2022

Rabroekenweg 20, 7942 JE Meppel  
P.O. Box 37, 7940 AA Meppel  
The Netherlands

T. +31 (0)522-26 87 00  
F. +31 (0)6-53 859 900  
E. boombv@boomlab.nl

[www.boomlab.nl](http://www.boomlab.nl)

Naam : Osmo HIGH  
Boom artikelnummer : 80902657  
Mechatronics artikelnummer : QRR030930 B01  
Batchnummer : 2302041B

Bewaarcondities : 4 - 30 °C, goed afgesloten  
Bereidingsdatum : 25 februari 2022  
Vervaldatum : 25 februari 2023

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	764 - 804	768
pH meting (37°C)	geen	7,35 – 7,45	7.41
viscositeit (37°C)	mPas·s	27 - 33	27.49
Aantal keren geautoclaveerd		max 1	1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.



