



RR Mechatronics Manufacturing B.V.

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAAG
The Netherlands

T +31 229 291 129
www.rrmechatronics.com

Certificate of analysis

Product name: **Osmo HIGH**
Article number: **QRR 030930**
LOT number: 2403191B
Expiry date: 2024-03-07

LOT release date: 2023-03-07

Test	Specification	Measured value
Osmolality (mOsm/kg)	764 - 804	799
pH at 25°C	7.35 - 7.45	7,43
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	30,32

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed:

Date:

13-mar-2023

QARA Department



ANALYSECERTIFICAAT

Afgifte, Meppel 8 maart 2023

Rabroekenweg 20, 7942 JE Meppel
P.O. Box 37, 7940 AA Meppel
The Netherlands

T. +31 (0)522-26 87 00
☎ +31 (0)6-53 859 900
E. boombv@boomlab.nl

www.boomlab.nl

Naam : Osmo HIGH
Boom artikelnummer : 80902657
Mechatronics artikelnummer : QRR030930 B01
Batchnummer : 2403191B

Bewaarcondities : 4 - 30 °C, goed afgesloten
Bereidingsdatum : 7 maart 2023
Vervaldatum : 7 maart 2024

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	764 - 804	799
pH meting (37°C)	geen	7,35 – 7,45	7.43
viscositeit (37°C)	mPas·s	27 - 33	30.32

Aantal keren geautoclaveerd : max 1 1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.



