

RR Mechatronics Manufacturing B.V.

P.O.Box 225  
1620 AE HOORN  
The Netherlands

De Corantijn 13  
1689 AN ZWAAG  
The Netherlands

T +31 229 291 129  
www.rrmechatronics.com

## Certificate of analysis

Product name: **Osmo HIGH**  
Article number: **QRR 030930**  
LOT number: 2403307B  
Expiry date: 2024-03-02

LOT release date: 2023-03-02

Test	Specification	Measured value
Osmolality (mOsm/kg)	764 - 804	776
pH at 25°C	7.35 - 7.45	7,39
Viscosity at 37.0 ± 0.5 °C (cP)	27.0 - 33.0	28,76

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Material Safety Data Sheet (MSDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at [gara@rrmechatronics.com](mailto:gara@rrmechatronics.com)

Signed: 

Date: 06. mar - 2023

QARA Department





# ANALYSECERTIFICAAT

Afgifte, Meppel 2 maart 2023

Rabroekenweg 20, 7942 JE Meppel  
P.O. Box 37, 7940 AA Meppel  
The Netherlands

T. +31 (0)522-26 87 00  
☎ +31 (0)6-53 859 900  
E. boombv@boomlab.nl

[www.boomlab.nl](http://www.boomlab.nl)

Naam : Osmo HIGH  
Boom artikelnummer : 80902657  
Mechatronics artikelnummer : QRR030930 B01  
Batchnummer : 2403307B

Bewaarcondities : 4 - 30 °C, goed afgesloten  
Bereidingsdatum : 02-03-2023  
Vervaldatum : 02-03-2024

<u>Omschrijving</u>	<u>Eenheid</u>	<u>Specificaties</u>	<u>Actuele batchwaarde</u>
osmolaliteit	mOsm/kg	764 - 804	776
pH meting (37°C)	geen	7,35 – 7,45	7,39
viscositeit (37°C)	mPas·s	27 - 33	28,76
Aantal keren geautoclaveerd		max 1	1

Akkoord : Afdeling Productie

Dit certificaat is een elektronisch gegenereerd document en geldig zonder handtekening.



