

## Final Inspection Protocol

<b>Type of sensor</b>	Integration of Oxygen Sensor PSt3
<b>Ordering code</b>	Int-SP-PSt3-YAU-D5-YOP
<b>Article number</b>	200000787
<b>Batch number</b>	200123-001_PSt3-1208-01
<b>Date</b>	17 Feb 2020
<b>Reference device</b>	SACF0002000523

**Dear customer,**

Data are specific for the test instrumentation used in our laboratory. Your instrument in combination with the sensors might show different signal amplitude and phase angles.

<b>Data</b>						
Atmospheric Pressure:	<input type="text" value="973"/>	hPa				
Calibration Mode	<input type="text" value="Humid"/>					
	Phase signal	Valid range	Temperature	Valid range	Amplitude	QC-passed?
	[°]	[°]	[C°]	[C°]	[µV]	(ok / failed)
cal 0 0 %a.s.	<b>60.94</b>	58.00 - 62.00	<b>19.8</b>	18.0 - 22.0	306592.4	OK
cal 2nd 100 %a.s.	<b>27.84</b>	25.00 - 29.00	<b>19.8</b>	18.0 - 22.0	129543.4	OK
<b>Response time [t90]:</b>	< 60 s	<b>Valid range:</b>		< 60 s		
	<input type="text"/>	<i>Please type in these values into the software for "manual calibration"</i>				

<b>Sensor Constants</b>		
<b>f1 = 0.810</b>	<b>dPhi1 = -0.06684</b>	<b>dKSV1 = 0.000394</b>
<b>m = 29.17</b>	<b>dPhi2 = -0.00035</b>	<b>dKSV2 = 0.000000</b>

*Sensor is within the accepted tolerance window*

*Sensor is visually inspected.*

Inspected by (name, date) Ch. Rapi 17. FEB. 2020

Approved by (name, date) Th. Eisenbeis 17. FEB. 2020