

EU DECLARATION OF CONFORMITY

RR Mechatronics Manufacturing B.V.

This declaration of conformity is issued under the sole responsibility of the manufacturer:

P.O.Box 225
1620 AE HOORN
The Netherlands

De Corantijn 13
1689 AN ZWAAG
The Netherlands

T +31 229 291 129

www.rrmechatronics.com

RR Mechatronics Manufacturing B.V.
De Corantijn 13, 1689 AN Zwaag, The Netherlands
SRN: NL-MF-000023105

We declare that:

Isotonic solution:

Trade name: **RPI-Check**

| | | | |
|-----------------|-------------|-------------------|----------------|
| Product models: | Model name: | Product-ID (REF): | Basic UDI-DI |
| | RPI-Check | A0020487 | 08719189137538 |

EMDN-code: W01030199 In vitro medical devices, reagents, heamatology / haemostasis, immunohaematology / histology / cytology, haematology reagents, haematology reagents - other

Classification IVDR: Class A

Intended purpose: RPI-Check is an isotonic solution that is used in the daily routine to monitor the operation of the measurement functions within the RPI. If a shift in performance is determined, it may indicate the RPI requires service

is in conformity with the requirements of the following EU legislation:

Regulation (EU) 2017/746 In vitro diagnostic medical devices
(conformity assessment according Article 48 of this regulation)

The following (harmonized) standards have been applied:

| | |
|---------------------|--|
| EN ISO 13485:2016 | Medical devices – Quality management systems – Requirements for regulatory purposes |
| EN 13612:2002 | Performance evaluation of in vitro diagnostic medical devices |
| EN ISO 14971:2012 | Medical devices – Application of risk management to medical devices |
| EN-ISO 15223-1:2016 | Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements |
| EN ISO 18113-1:2011 | In vitro medical devices – Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements |
| EN ISO 18113-2:2011 | In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 2: In vitro diagnostic reagents for professional use |
| EN ISO 23640:2015 | In vitro medical devices – Evaluation of stability of in vitro diagnostic reagents |

The CE mark was applied for the first time on this type of IVD reagent in 2021.

Place: Zwaag, The Netherlands

Date: May 10, 2022

Signature:

Name: Jan Buis

Function title: CEO
RR Mechatronics Manufacturing B.V.

Stamp:

