

PACKAGE INSERT**REF** QRR010905**Supplier**

Mechatronics Instruments B.V.
P.O. Box 225
1620 AE Hoorn
The Netherlands
Phone: +31 229 291 129
Fax: +31 229 241 534

Starrsed Cleaning Agent**Intended purpose**

Product to remove protein deposition from the Westergren pipettes of Starrsed ESR analyzers.

Instructions For Use

The product is for use on the Starrsed analysers only. The product is only to be used in a dilution, refer to the Starrsed instrument manual for detailed instructions.

1. Fill the clean adapter with hot de-ionized water till the first mark in the adapter.
2. Add Starrsed Cleaning Agent **till the second marker in the adapter.**
3. Place the cap(s) on the adapter and mix well.
4. Place the filled clean adapter on the instrument according the instrument instructions and start the Fill and Clean procedure.

Chemical/Physical data

Contains surfactants in an alkaline aqueous environment as active ingredient. To booster the cleaning performance, this reagent contains some excipients (builder and sequestering agent).

The latest version of the Safety Data Sheet is available on the website www.rrmechatronics.com

Transport/Storage/Packaging

Stability during transport: 5 - 60 °C, with a maximum of 14 days.

Store the concentrate at 15 - 25 °C, in the container in which it is supplied, sealed, out of direct sunlight and away from strong acids and aluminium.

Starrsed Cleaning agent is supplied in 100 ml bottles. QRR010905: 1 bottle

Safety

Any serious incident that has occurred in relation to this product must be reported to the distributor and/or manufacturer and the competent local authorities.

Shelf life

After production: 5 years, but no longer than the expiry date.

In-use stability: Maximum of 3 months at advised temperature of 5 till 25°C.

Ecological information

Product is biodegradable.

Disposal

Recommendation Must not be disposed of together with household garbage. Do not allow product to reach sewage system. Disposal must be made according to official regulations.

Recommended cleaning agent Water, if necessary with cleaning agent.



RR Mechatronics Manufacturing B.V.
De Corantijn 13, 1689 AN ZWaaag, The Netherlands
www.rrmechatronics.com

Rev. 5

Revision due to IVDR-registration, intended purpose description modified

Website added under manufacturer address

Component description and safety chapter added

Storage temperature changed (equal to label)

Order unit change from 2 to 1