



Work instruction Number 234

Page 1 of 1

Purpose: Maintenance level 3

Safety: *Bio Hazard area*

Instrument: Starrsed TL

Revision: 002, 2019

- Clean fill nozzle and exchange O-ring Fill Nozzle
- Clean liquid separator
- Exchange peristaltic pump tubes
- Fill & Clean
- Check/Clean liquid sensor
- Sensor check
- Exchange waste pump tube
- Exchange pinch valve
- Final cleaning and preparation
- Air filter replacement